

EXIT

Heart Rate
HR Source Smart
Heart Rate Tone Vol OFF

NIBP
NIBP tone NONE

ECG
Cable 3 lead
Pace detect OFF
Filter Monitor
Sensitivity MEDIUM

Temperature OFF
Unit of measure °F

Respiration ON

SpO2
Average 12 seconds
Search time 20 seconds

CO2 ON
Unit of measure mmHg
N2O compensation OFF
CO2 Waveform Fill YES

PARAMS

<u>EXIT</u>	TYPE	GAIN	SWEEP MM/SEC	SIZE
Waveform 1	ECG II	x1.0	25.0	25mm
Waveform 2	PLETH	x2.0	25.0	25mm
Waveform 3	ET CO2	x1.0	12.5	25mm
Waveform 4	MINI-TREND	x1.0	25.0	25mm
Waveform 5	OFF	x1.0	25.0	25mm
Waveform 6	OFF	x1.0	25.0	12mm
External Display		OFF		
Display Lines		OFF		

DISPLAY

EXIT

Print Type	Graphical
Alarm Print	OFF
BP Print	OFF
Interval Print	OFF
Interval Print Type	Tabular
Snapshot Size	6 seconds
History Size	6 seconds
Waveform 1	ECG II
Gain	x1.0
Waveform 2	OFF
Gain	x1.0
Printer Speed	25.0 MM/SEC
Serial Type	Tabular
Serial	Format CUSP
	Baudrate 38400

PRINT